DECLARATION OF CONFORMITY

IN ACCORDANCE WITH ISO/IEC 17050:2004



EMC

EN 55022:1998+A1:2000+A2:2003 IEC 61326-1:1997+A1:1998+A2:2001/EN 61326-1:1997+A1:1998+A2:2001 IEC 61000-4-2:2001 / EN 61000-4-2:1995+A1:1998+A2:2001 IEC 61000-4-3:1995+A1:1998+A2:2001 EN 61000-4-3:1996+A1:1998+A2:2001

ICES-001: 1998 (Canada) CFR 47 FCC Part 15, Subpart B (Exempted devices) (USA) Limits applicable to Class A equipment

GFIS

-IER

4kV CD, 8kV AD

Limit

3V/m, 80-1000MHz

Supplemental Information:

The product was tested in a typical configuration with Kingfisher International test systems.

9 September 2016

Date

Bruce Robertson

Brue Rotate

Name Director Title

For further information, please contact your local Kingfisher International sales office, agent or distributor.