



**DECLARATION OF CONFORMITY**  
IN ACCORDANCE WITH ISO/IEC 17050:2004



**Manufacturer's Name:** Kingfisher International Pty. Ltd.  
**Manufacturer's Address:** 720 Springvale Road, Mulgrave, Victoria 3170, Australia

herby declares, that the products listed below

<b>Product Name:</b>	<b>Tone and Traffic Identifier</b>
<b>Model Number:</b>	KI 6171
<b>Product Options:</b>	N/A

comply with the essential requirements of the applicable European Directives:

- Low Voltage Directive 73/23/EEC and the EMC Directive 89/336/EEC, amended by 93/68/EEC, and carries the CE marking accordingly.
- Directive 2002/95/EC on restriction of the use of certain hazardous substances in electrical and electronic equipment (RoHS).
- Directive 2002/96/EC on waste electrical and electronic equipment (WEEE). The instrument can be disposed of local retail shop with WEEE collection point.

and conform to the following standards and specifications:

MIL-PRF-28800F: 1996 Performance specification-Test equipment for use with electrical and electronic equipment, general specification  
IEC 60529: 2001/ EN 60529:1993+A1:2003 Degrees of protection provided by enclosures

**EMC**

EN 55022:1998+A1:2000+A2:2003  
IEC 61326-1:1997+ A1:1998+A2:2001/EN 61326-1:1997+A1:1998+A2:2001  
IEC 61000-4-2:2001 / EN 61000-4-2:1995+A1:1998+A2:2001  
IEC 61000-4-3:1995 +A1:1998+A2:2000/  
EN 61000-4-3:1996+A1:1998+A2:2001

**Limit**


Limits applicable to Class A equipment  
  
4kV CD, 8kV AD  
  
3V/m, 80-1000MHz

ICES-001: 1998 (Canada)  
CFR 47 FCC Part 15, Subpart B (Exempted devices) (USA)

**Supplemental Information:**

*The product was tested in a typical configuration with Kingfisher International test systems.*

9 September 2016  
Date

Bruce Robertson  
Name  
  
Director  
Title

For further information, please contact your local Kingfisher International sales office, agent or distributor.