

DECLARATION OF CONFORMITY IN ACCORDANCE WITH ISO/IFC 17050:2004



Manufacturer's Name: Kingfisher International Pty. Ltd.

Manufacturer's Address: 720 Springvale Road, Mulgrave, Victoria 3170, Australia

hereby declares, that the products listed below

Product Name: Optical CWDM Power Meter Model Number: KI6501 / T6501 Series

Product Options: This declaration covers all options of the above product(s)

complies with the essential requirements of the applicable European Directives:

- Low Voltage Directive 2014/35/UE and the EMC Directive 2014/30/UE, and carries the CE marking accordingly.
- Directive 2011/65/EU on restriction of the use of certain hazardous substances in electrical and electronic equipment (RoHS).
- Directive 2012/19/EU on waste electrical and electronic equipment (WEEE). The instrument can be disposed of local retail shop with WEEE
 collection point.

and conforms to the following standards and specifications:

MIL-PRF-28800F:1996 Performance specification -Test equipment for use with electrical and electronic equipment, general specification IEC 60529 Ed. 2.2 / EM 60529:1991 Degrees of protection provided by enclosures

IEC 61315 Ed. 2.0 Calibration of fibre-optic power meters

EMC

IEC 61326-1 Ed. 2.0 / EN 61326-1:2013 IEC CISPR 11 Ed. 6.1 / EN 55011:2016 IEC CISPR 16-1:2004~2019 IEC CISPR 16-2:2003~2016

IEC 61000-4-2 Ed. 2.0 (2008) / EN 61000-4-2:2009 IEC 61000-4-3 Ed. 3.2 / EN 61000-4-3:2006

ICES-001: 2006 (Canada)

CFR 47 FCC Part 15, Subpart B (Exempted devices) (USA)

Limit

3V/m. 80-1000MHz

Limits applicable to Group 1 Class B equipment Limits applicable to Class B equipment Limits applicable to Class B equipment 4kV CD, 8kV AD

Supplemental Information:

The product was tested in a typical configuration with Kingfisher International test systems.

17 Oct 2019	Bruce Robertson
Date	Name
	Director
	Title

For further information, please contact your local Kingfisher International sales office, agent or distributor.

Revision: D Issue Date: 17 Oct 2019